

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130492</u> ✓
Name <u>Shayne Calt</u>		ID# <u>4559</u> Date <u>11-28-08</u> ✓
A	Agency <u>AST</u>	Phone # <u>733-2256</u>
Instrument Location <u>Willow AST</u> ✓		
Alco S/N <u>75470</u> ✓ Target Value <u>.076</u> ✓ High Pressure <u>800</u> ✓		
Alco Test Value Average <u>.076</u> ✓ <u>.075</u> ✓		
1 st Alco 2 nd Alco		
Signature <u>[Signature]</u> ✓		
COB 12/5/08		

Under the Alaska Rules of Evidence, I certify that: (OVER)

(Do Not write in the area below)
(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Shayne M Calt, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/5/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130492 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

NOVEMBER 28, 2008

OPERATOR'S NAME:

CALT SHAYNE M

OPERATOR'S NUMBER: 4559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

V

O.L. #: V

DEPT/AGENCY: TALO

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .076 ✓

ALCO S/N: 75470

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.13 in

ALCO TARGET .074 20:38

BLANK TEST .000 20:39

INTERNAL STANDARD VERIFIED 20:39

ALCO TV 29.13 in .076 ✓ 20:39

BLANK TEST .000 20:40

SUBJECT SAMPLE .000 20:41

BLANK TEST .000 20:41

ALCO TV 29.13 in .075 ✓ 20:42

BLANK TEST .000 20:43

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

NOVEMBER 28, 2008 ✓

TIME 20:36

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.13 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~